



Urgent Field Safety Notice

Mechanical Bileaflet Prosthetic Valves 02/27/2018 Device Redesign

Date: 02/27/2018

Attention: Cardiothoracic/Cardiovascular surgeons, Risk Managers, Medical/Surgical Directors, and OR Materials Managers

Details on affected devices:

<u>Brand Name</u>	<u>Product Code</u>	<u>Lot Number</u>
On-X Prosthetic Heart Valves	ONXA, ONXAE, ONXAC, ONXACE, ONXM, ONXMC, ONXAN, ONXANE	All On-X Aortic and Mitral Heart Valves

Description of the problem:

The Medicines & Healthcare products Regulatory Agency (MHRA) issued a Medical Device Alert for Replacement bileaflet mechanical heart valves – risk of inverted implantation. This alert was initiated by MHRA based on MHRA becoming aware of five (5) incidents worldwide over the last fifteen (15) years where mechanical bileaflet prosthetic valves were reported to have been implanted in an inverted position because they were put on their holder upside down. To date, On-X Life Technologies, Inc. has received one (1) related report. The probability of this error remains very low relative to the number of valves successfully implanted during this time.

Action taken by On-X Life Technologies

Although the risk of inverted valve implantation cannot be completely eliminated, in an effort to minimize future events, On-X Life Technologies, Inc. has initiated a redesign of the valve holder to be unidirectional; e.g., it can only be used in the correct position as identified in the instructions for use. On-X unidirectional holders are unlikely to become available for at least a year. The current valve holders will continue to be available until re-design efforts are completed and all required regulatory clearance notifications have been received. In the interim, it is important that clinical personnel handling the On-X heart valve are aware of the serious risk associated with the improper use and implant of an inverted valve.

Transmission of this Field Safety Notice and Actions Required:

This notice needs to be passed on to all those who need to be aware within your organization. The following actions are required:

- Read and understand this Safety Alert and distribute as necessary to all relevant personnel.
- Strictly follow the On-X Valve Instruction for Use (IFU) that accompanies each device, which states: WARNING: DO NOT attempt to reinsert the valve holder into the valve once it has been removed.
- Ensure the surgical team is trained on the proper handling and implant of the On-X valve, including awareness of all warnings and precautions stated in the IFU.



CryoLife[®]
Life Restoring Technologies[®]

Contact Address:

fieldassurance@cryolife.com

The undersign confirms that this notice has been sent to the appropriate Regulatory Agency.

On-X Life Technologies, Inc. thanks you for your attention to this Safety Alert. If you have any questions or require further information, please contact the email address listed above.

Tamara Cardona
Manager, Field Assurance
CryoLife, Inc.