

Drägerwerk AG & Co. KGaA, 23542 Lübeck

To our customers with the Dräger Oxylog VE300 ventilator with carrying system holder

April, 2019

Important safety notice!!!

**The Oxylog VE300 may fall out of the carrying system holder.
Only the carrying system holder is affected, other devices are NOT affected!**

Dear Sir or Madam,

During the course of our continuous market surveillance, a total of two cases have been made known to us in which the Oxylog VE300 ventilator had detached from or was very easy to detach from the carrying system holder on the wall. The cause of this is a spring of the carrying system holder not being tensioned to the specified degree. This can lead to the device coming loose due to external forces, especially during mobile use, and, in the worst case, falling on either a patient or user. We are as yet unaware of any cases of patient or user injury due to this. A risk of injury cannot, however, be excluded due to the weight of the device.

According to our records, you have received one or more carrying system holders from us.


We are sending you a new, error-free carrying system holder together with this safety notice free of charge.

Error-free carrying system holders can be identified by a small black marking that is visible when the grey cover is removed (blue circle in the illustration).

Please remove your existing carrying system holder.

Please install the new carrying system holder and check for proper installation in accordance with the assembly instructions.

Please send the faulty carrying system holder back to Dräger and confirm the new installation by including the completed enclosed response form in the return delivery.



Drägerwerk AG & Co. KGaA
Moislinger Allee 53-55
23558 Lübeck, Germany
Postal address:
23542 Lübeck, Germany
Tel. +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
VAT ID: DE135082211

Bank information:
Commerzbank AG, Lübeck
IBAN: DE95 2304 0022 0014 6795 00
SWIFT code: COBA DE FF 230
Sparkasse zu Lübeck
IBAN: DE15 2305 0101 0001 0711 17
SWIFT code: NOLADE21SPL

Company headquarters: Lübeck
Commercial register:
Municipal Court Lübeck HRB 7903 HL
General partner: Drägerwerk Verwaltungs
AG
Company headquarters: Lübeck
Commercial register:
Municipal Court Lübeck HRB 7395 HL

Chairman of the Executive Board of
Drägerwerk AG & Co. KGaA and
Drägerwerk Verwaltungs AG:
Prof. Dr. Nikolaus Schweickart
Board of Directors:
Stefan Dräger (Chairman)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner

We have informed the competent authorities of this measure.

We apologise for any inconvenience this may cause you. If you have any further questions, please contact your local Dräger representative.

With kind regards,

A handwritten signature in blue ink, appearing to read 'i.v. Delf', is positioned above the printed name and title.

Michael Grossmann
Head of Product Management
Respiratory Care

Attached:

- Customer response form
- New carrying system holder