

WEINMANN Emergency Medical Technology GmbH + Co. KG  
PO Box 57 01 53 • 22770 Hamburg • GERMANY

Hamburg, January 2023

## **Important safety information: Field safety corrective action on a medical device**

**Reference:** FSCA MMS2 2023-01.01

**From:**

WEINMANN Emergency Medical Technology GmbH + Co. KG

**Addressee:**

Users and operators, specialist dealer partners

**Medical devices affected** (trade name and article no. of products):

- MEDUMAT Standard<sup>2</sup>; WM 28710-01, WM 28710-02, WM 28710-03 and WM 28710-04 up to and including serial number SN 19645.

Dear Customers,

Quality and safety are our top priority, which is why we wish to act in a consistent and transparent manner as usual and, in the context of your obligation to co-operate under medical devices legislation, ask you to implement this corrective action so that users can continue to use our products on patients safely.

You may continue using your MEDUMAT Standard<sup>2</sup> until the remedial measure described below has been completed.

### **1. Problem description and cause**

We have noticed during our regular internal quality checks that in rare cases a device start as part of the device self-test may lead to a device malfunction, rendering the device not ready for use straightaway. Performing a restart allows the device self-test to be passed. The cause relates to the faulty communication of the internal differential pressure sensor.

Page 1 of 4

**Company Headquarters**  
WEINMANN Emergency  
Medical Technology GmbH + Co. KG  
Frohbösestraße 12 • 22525 Hamburg • GERMANY  
T: +49 40 88 18 96-0  
F: +49 40 88 18 96-480  
www.weinmann-emergency.com

**Center for Production, Logistics, Service**  
WEINMANN Emergency  
Medical Technology GmbH + Co. KG  
Siebenstücken 14 • 24558 Henstedt-Ulzburg  
GERMANY

**Business Management**  
Dipl.-Volksw. Marc Griefahn  
Dipl.-Kfm. Philipp Schroeder  
Dipl.-Volksw. André Schulte

**Registration Court**  
Hamburg Municipal Court  
Dept. A # 115967  
V.A.T. # DE288367727  
WEEE Reg. # DE 47913245

**Creditor ID**  
DE35ZZZ00000353971

**General Partner**  
WEINMANN Emergency  
Management GmbH, Hamburg

**Registration Court**  
Hamburg Municipal Court  
Dept. B # 38144

**Certified QM System meeting**  
EC directive 93/42/EEC, Annex II  
(EN ISO 9001/EN ISO 13485)

**Banking Connections**

**Deutsche Bank AG Hamburg**  
IBAN DE87 2007 0000 0646 9639 00  
SWIFT DEUTDEHH

**Hamburger Sparkasse AG**  
IBAN DE44 2005 0550 1032 2626 67  
SWIFT HASPDEHHXXX

**Commerzbank AG Hamburg**  
IBAN DE14 2004 0000 0632 0071 00  
SWIFT COBADEHHXXX

## 2. What is the risk to the patient?

In individual cases, the above fault may make therapy impossible or delay therapy. In this case, an alternative means of ventilation must be used.

## 3. Remedy

The remedy consists of updating to software version 5.5. This optimizes the device start process and the sensor communication. In rare cases, the start process may take slightly longer to complete. Once the device self-test has been passed, the device is ready for use as normal.

## 4. What measures should the addressee take?

This letter contains a report form "Reporting regarding safety information".

Please take the following measures as soon as possible:

- Use the attached report form to **confirm receipt of this letter** no later than **01/31/2023**.
- Ensure that this safety information is brought to the attention of all users of the above-mentioned product and any other people in your organization that need to be informed.
- If you have already resold the products, please forward a copy of this information on to your customers.
- Download the new software version 5.5 for MEDUMAT Standard<sup>2</sup>. The update files are available to download from our download page ([www.weinmann-emergency.com/sw-update-55-mms2](http://www.weinmann-emergency.com/sw-update-55-mms2)) (software package: MEDUMAT\_Standard2\_SW\_5.5.zip).
- If you do not currently have the instructions for use for SW version 5.1 or higher, these are available in the Download Center ([www.weinmann-emergency.com/download/downloadcenter/](http://www.weinmann-emergency.com/download/downloadcenter/)) or you can request instructions for use from us via the following link ([www.weinmann-emergency.com/download/form-request-instructions-for-use/](http://www.weinmann-emergency.com/download/form-request-instructions-for-use/)) using the order form available there.
- Install software version 5.5 on all your devices. Section 4 "Updating software" of the instructions for use for MEDUMAT Standard<sup>2</sup> includes details on how to update the software.
- Report to us that you have updated your software for the specific device. To do so, use the online form on the software download page. If this is not possible, please use the documentation form included in the MEDUMAT\_Standard2\_SW\_5.5.zip software package as an alternative.
- Please perform all **corrective action by no later than 02/28/2023**.

As mentioned above, you may continue using your MEDUMAT Standard<sup>2</sup> until the remedial measure described has been completed. However, please ensure that an alternative means of ventilation (e.g., bag-valve mask) is available.

This corrective action is mandatory. The responsible authority has been informed of the procedure.

**Contact**

If you have any questions or need support, please contact your local specialist dealer or contact us directly:

Phone: +49 40 88 18 96 – 122


e-mail: [AfterSalesService@weinmann-emt.de](mailto:AfterSalesService@weinmann-emt.de).

Kind regards,

WEINMANN Emergency  
Medical Technology GmbH + Co. KG



André Schulte  
Managing Director



ppa. Dennis Horstmann  
Authorized Signatory  
Head of Supply Chain + Quality Management

**Enclosures**

Form: "Reporting regarding safety information"

# Report to WEINMANN Emergency by 2023-01-31

Regarding MEDUMAT Standard<sup>2</sup> safety information: Reference: FSCA MMS2 2023-01.01

Please fill in this report form in full and return it by e-mail, fax or mail to:

E-mail: **AfterSalesService@weinmann-emt.de**  
Fax: **+49 40 88 18 96 - 490**

**WEINMANN Emergency Medical Technology GmbH + Co. KG**  
Technischer Service  
Frohösestraße 12  
22525 Hamburg, GERMANY

I hereby confirm receipt of this letter and that I have read, understood and will implement its contents. This letter has been brought to the attention of all users of the product and of other people in my organization who need to be informed.

If the products have been passed on to third parties (applies to specialist dealers, for example), a copy of this information has been passed on to them.

Please complete in full in block capitals:

Company/organization details:

Customer no.:

\_\_\_\_\_

Company/organization + address:

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

I am no longer in possession of the medical device:

The device has been scrapped

The new owner is (company + address)

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_  
Date, signature

\_\_\_\_\_  
Name (in block letters)

\_\_\_\_\_  
Position (in block letters)

\_\_\_\_\_  
e-mail address (in block letters)