

Siemens Healthcare Diagnostics GmbH, HC CEMEA CEE QT, Siemensstrasse 90,
1210 Vienna

Name	M.A. Roland Ertl
Department	HC CEMEA CEE QT
Telephone	+43 51707-38274
Mobile	+43 (664) 8011738274
E-mail	roland.re.ertl@siemens.com
Date	Decemver 22 nd 2017
Document Ref#	IMC18-02.A.OUS

Urgent Field Safety Notice:

IMMULITE®/IMMULITE® 1000 IMMULITE® 2000/IMMULITE® 2000 XPi Assays

Biotin Interference

Dear Sirs,

Our records indicate that your facility may have received the products listed in Table 1 below.

Table 1. Affected Products-All lots

Assay	Catalog Number	Siemens Material Number (SMN)
3gAllergy™ Specific IgE Universal Kit IMMULITE 2000/IMMULITE 2000 XPi	L2KUN6 L2KUNJ6	10380875 10711939
Anti Hbc IMMULITE 2000/IMMULITE 2000 XPi	L2KHC2	10381311
BR-MA (CA15-3) IMMULITE/IMMULITE 1000	LKBR1	10380948
BR-MA (CA15-3) IMMULITE 2000/IMMULITE 2000 XPi	L2KBR2	10380983
CEA IMMULITE/IMMULITE 1000	LKCE1	10380945
CEA IMMULITE 2000/IMMULITE 2000 XPi	L2KCE2 L2KCE6	10380994 10380995
CK-MB IMMULITE/IMMULITE 1000	LKMB1	10381016
CK-MB IMMULITE 2000/IMMULITE 2000 XPi	L2KMB2	10381033
EPO IMMULITE/IMMULITE 1000	LKEPN1	10487627
EPO IMMULITE 2000/IMMULITE 2000 XPi	L2KEPN2 L2KEPN6	10487628 10487629

Siemens Healthcare Diagnostics GmbH
Management: Wolfgang Koepl, Stefan Scheidler

Siemensstrasse 90
1210 Vienna
Austria

Tel.: +43 51707 0
healthcare.siemens.com

Assay	Catalog Number	Siemens Material Number (SMN)
Folic Acid IMMULITE 2000/IMMULITE 2000 XPi	L2KFO2 L2KFO6	10380911 10380912
Gastrin IMMULITE/IMMULITE 1000	LKGA1	10380962
Gastrin IMMULITE 2000/IMMULITE 2000 XPi	L2KGA2	10380979
OM-MA (CA125) IMMULITE/IMMULITE 1000	LKOP1	10380969
OM-MA (CA125) IMMULITE 2000/IMMULITE 2000 XPi	L2KOP2	10380972
Thyroglobulin IMMULITE/IMMULITE 1000	LKTY1	10381644
Thyroglobulin IMMULITE 2000/IMMULITE 2000 XPi	L2KTY2	10381648
Vitamin B12 IMMULITE/IMMULITE 1000	LKVB1	10380900

Reason for Correction

Siemens Healthcare Diagnostics has confirmed through internal investigation that the IMMULITE[®]/IMMULITE[®] 1000/IMMULITE[®] 2000/IMMULITE[®] 2000 XPi assays listed in Table 1 are susceptible to Biotin interference. This occurs when biotin present in patient samples interferes with the biotin-streptavidin assay architecture on the IMMULITE platform. Biotin interference has the potential to bias analytical results on the assays listed above. The Instructions for Use (IFU) currently do not list biotin as a potential interferant.

Concentrations of biotin above the concentration listed in Table 2 can potentially result in interference greater than 10%, leading to either falsely elevated or falsely depressed results.

CEA, Folic Acid, OM-MA (CA125) and Vitamin B12 exhibited falsely elevated results with biotin concentrations above those listed in Table 2.

3gAllergy Specific IgE, Anti HBc, BR-MA (CA15-3), CK-MB, EPO, Gastrin and Thyroglobulin exhibited falsely depressed results with biotin concentrations above those listed in Table 2.

Table 2. Biotin concentration at which less than or equal to 10% bias was observed

Assay	Catalog Number	Siemens Material Number (SMN)	Biotin Concentration ng/mL [nmol/L]*
3gAllergy™ Specific IgE Universal Kit IMMULITE 2000/IMMULITE 2000 XPi	L2KUN6 L2KUNJ6	10380875 10711939	5 [20]
Anti HBc IMMULITE 2000/IMMULITE 2000 XPi	L2KHC2	10381311	5 [20]
BR-MA (CA15-3) IMMULITE/IMMULITE 1000	LKBR1	10380948	100 [409]
BR-MA (CA15-3) IMMULITE 2000/IMMULITE 2000 XPi	L2KBR2	10380983	100 [409]
CEA IMMULITE/IMMULITE 1000	LKCE1	10380945	9 [37]
CEA IMMULITE 2000/IMMULITE 2000 XPi	L2KCE2 L2KCE6	10380994 10380995	2 [8]
CK-MB IMMULITE/IMMULITE 1000	LKMB1	10381016	2 [8]
CK-MB IMMULITE 2000/IMMULITE 2000 XPi	L2KMB2	10381033	5 [20]
EPO IMMULITE/IMMULITE 1000	LKEPN1	10487627	5 [20]
EPO IMMULITE 2000/IMMULITE 2000 XPi	L2KEPN2 L2KEPN6	10487628 10487629	2 [8]
Folic Acid IMMULITE 2000/IMMULITE 2000 XPi	L2KFO2 L2KFO6	10380911 10380912	500 [2047]
Gastrin IMMULITE/IMMULITE 1000	LKGA1	10380962	2 [8]
Gastrin IMMULITE 2000/IMMULITE 2000 XPi	L2KGA2	10380979	2 [8]
OM-MA (CA125) IMMULITE/IMMULITE 1000	LKOP1	10380969	2 [8]
OM-MA (CA125) IMMULITE 2000/IMMULITE 2000 XPi	L2KOP2	10380972	5 [20]
Thyroglobulin IMMULITE/IMMULITE 1000	LKTY1	10381644	5 [20]
Thyroglobulin IMMULITE 2000/IMMULITE 2000 XPi	L2KTY2	10381648	5 [20]
Vitamin B12 IMMULITE/IMMULITE 1000	LKVB1	10380900	250 [1023]

* Concentrations of biotin above the concentration listed can potentially result in interference greater than 10%.

Risk to Health

The probability of misinterpretation of results for the above assays due to this issue is remote. Mitigations include correlation to clinical history and presentation as well as to other diagnostic laboratory testing, serial testing, and/or concomitant imaging studies depending on the analyte. Siemens is not recommending a lookback as a result of this issue.

Actions to be Taken by the Customer

- Please refer to the information provided in Table 2 until the appropriate IFU updates regarding biotin interference are completed.
- Review this letter with your Medical Director.
- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- If you have received any complaints of illness or adverse events associated with the products listed in Table 1, immediately contact your local Siemens Customer Care Center or your local Siemens Technical Support Representative.

Please retain this letter with your laboratory records, and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Customer Care Center or your local Siemens Technical Support Representative.

Sincerely yours,

Siemens Healthcare Diagnostics GmbH



i.V. Dipl. Ing. Franz Schwarz
Head of RAQS Austria & SEE



i.A. Martina Sattler, MSc., MSc.
Product Manager Austria & SEE