

IMPORTANT FIELD SAFETY NOTIFICATION

Subject: Incorrect Monitor Unit Scaling Using a Specific Monaco[®] Workflow

Product: Monaco[®]

Scope: Sites affected will be those who:

- Created 3D plans using Monaco[®] version 5.10 or 5.11, and
- Used Elekta motorized wedges

Notification Released: June, 2019

Description of Problem:

Monitor Units can be incorrect when following a specific workflow in Monaco[®].

Details:

When creating 3D plans using either MU or Dose weighting modes, if the user changes the number of fractions, rescales the plan and then changes the wedge angle, the Monitor Units are scaled incorrectly.

Clinical Impact:

If the Monitor Units are not correct, the patient will be incorrectly treated. There could be a critical overdose or under-dose proportional to the Rescale change.

Recommended User Action:

Prior to treatment, independent dose calculation checks and secondary MU checks should always be done. Both should be standards of practice in radiation therapy clinics and will detect the problem.

The problem can be avoided by forcing a Monaco[®] recalculation (change dose calculation grid spacing and change back) when any wedge angle change is made.

IMPORTANT FIELD SAFETY NOTIFICATION

This document contains important information for the continued safe and proper use of your equipment.

- Please post this notice in a place accessible to all users, e.g. Instructions for Use, until this action is closed.
- Advise the appropriate personnel, working with this product, on the content of this letter.

Elekta Corrective Actions:

Your site will be notified when a software fix is available. Note that sites running Monaco[®] versions 5.40 and above are not affected.

This notice has been submitted to the appropriate Regulatory Authorities.

We sincerely apologize for any inconvenience this action may cause and thank you in advance for your cooperation.

IMPORTANT FIELD SAFETY NOTIFICATION

Acknowledgement Form

In order to meet regulatory requirements, you are required to complete this form and return it to Elekta immediately upon receipt, but no later than within 30 days.

Classification: Important Field Safety Notification	FCO Reference Number: 382-01-MON-013
Description Incorrect Monitor Unit Scaling Using a Specific Monaco [®] Workflow	

Hospital:	
Device Serial No(s): (if applicable)	Location or Site:

I acknowledge that I have read and understood this Notice and accept the implementation of any given recommendation.	
Name:	Title:
Customer Signature:	Date:

New installation confirmation to be signed by the installing Elekta engineer or a Representative employee, when the installed product has a physical IFU/manual:	
I acknowledge that the customer has been informed on the content of this notice and that it has been inserted into the applicable copy of the User Manual, or added on record with the applicable User Manual:	
Name:	Title:
Signature:	Date: