

Urgent Field Safety Notice

New Calibrator Values for ACID PHOSPHATASE, ALKALINE PHOSPHATASE opt. liquicolor, LDH SCE mod. liquiUV and PHOSPHORUS liquirapid in AUTOCAL and re-assigned Target Values in HUMAN Controls Advice

November 09, 2023

Attention:

Distributors of HUMAN and end users applying the above-mentioned products:

Details on affected reagent, standard and control LOTs:

Product Name	REFs	LOTs
AUTOCAL	13160	0018 including all sublots up to 0018P, 0019 including all sublots up to 0019C
HumaTrol N	13511	0007
HumaTrol P	13512	0005
SERODOS®	13951	0006
SERODOS® PLUS	13151	0007, 0008 including all sublots up to 0008C
ACID PHOSPHATASE	10660, 12660, 10660600	All
ALKALINE PHOSPHATASE opt. liquicolor	12017, 12027, 12037, 12217, 12027600	All
LDH SCE mod. liquiUV	12014, 12024, 12214, 12014600, 12014300	All
PHOSPHORUS liquirapid	10027, 10027600, 10027300	All

Description of the problem:

In the course of our quality assurance activities, we noted that the current calibrator values for ACID PHOSPHATASE, ALKALINE PHOSPHATASE opt. liquicolor, LDH SCE mod. liquiUV and PHOSPHORUS liquirapid in AUTOCAL LOT 0018 are falsely elevated up to 17.9%. These deviating calibrator values are caused by decreasing enzyme activities and altered (internal) traceability procedures. Therefore, the calibrator values for certain analytes in AUTOCAL LOT 0018 had to be revised as shown in the table below:

AUTOCAL LOT 0018	Old Calibrator values Version 4/07-2023	Unit	Revised Calibrator values Version 5/09-2023	deviation old vs. revised AUTOCAL values
ACID PHOSPHATASE	16.3	U/l	14.6	11.6 %
	0.27	µkat/l	0.24	
ALKALINE PHOSPHATASE opt. liquicolor	407	U/l	348	17.0 %
	6.78	µkat/l	5.8	
LDH SCE mod. liquiUV	666	U/l	565	17.9 %
	11.1	µkat/l	9.41	
PHOSPHORUS liquirapid	7.43	mg/dl	6.86	8.3 %
	2.40	mmol/l	2.21	

Calibration values in AUTOCAL LOT 0019 are not affected, values released by version 2/07-2023 are still valid.

Although the above-indicated falsely elevated value would not directly harm the patient, results should nevertheless be checked, if unexpected or unplausible concentrations have been measured.

As a consequence the re-assigned calibrator values in AUTOCAL LOT 0018 are harmonized with those of the current AUTOCAL LOT 0019. In this context control values as well as control ranges in all above-mentioned control LOTs have been revised (marked in grey in the target value sheets) after re-evaluation and need to be applied for quality control of each specific analyte.

Please refer to the newly assigned target values and control ranges attached to this customer notification which also can be found on our website under the following link:

<https://www.human.de/clinical-chemistry/target-value-sheets>

or QR code:



Or in the log-in area for distributors: <https://www.human.de/login>

Advice on action to be taken by:

Distributor:

Please inform your customers about the issue of the affected Lots, based on this Urgent Field Safety Notice.

Please fill in the attached Reply Form confirming receipt of this Urgent Field Safety Notice and send it to support@human.de.

User:

Please ensure that the instructions resulting from this Urgent Field Safety Notice are implemented in the laboratory accordingly.

End users should confirm receipt of this Urgent Field Safety Notice to the local distributor.

Transmission of this Urgent Field Safety Notice:

This notice needs to be passed on all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred.

Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of this corrective action.

The Federal Institute for Drugs and Medical Devices (BfArM) and National Competent Authorities of European countries which are affected by the recall have received a copy of this Urgent Field Safety Notice.


Contact reference person:

(For distributors only. Distributors should provide their own detailed contact information to their end users):

Schuh, Jasmin
e-mail: support@human.de
Telephone: +49-6122-9988-333

We regret the inconvenience.

With kind regards,



Schuh, Jasmin
Customer Support & Applications



Dr. Märker-Stilz, Tina
Product Manager

Attachment
Reply Form

Reply Form

Urgent Field Safety Notice

**New Calibrator Values for ACID PHOSPHATASE, ALKALINE PHOSPHATASE opt. liquicolor, LDH SCE mod. liquiUV and PHOSPHORUS liquirapid in AUTOCAL and re-assigned Target Values in HUMAN Controls
Advice**

Please return by e-mail this filled in and signed Reply Form latest until November 23, 2023 to:

support@human.de

I confirm receipt of this Urgent Field Safety Notice and have informed all end users, who have obtained the affected Lots, in writing about the problem and the HUMAN recommendations.

If requested by national regulations, I have informed the respective authorities about the problem. (Note: to comply with European regulatory requirements HUMAN will inform European competent authorities directly.)

For distributors in the European Economic Area (EEA) and Turkey:
Please also provide the Urgent Field Safety Notice in your national language, which you have sent out to your end customers, as HUMAN will be approached by your national competent authority to provide this.

Date: _____

Company: _____

Name: _____

Signature: _____