



DiaMed GmbH  
Pra Rond 23  
1785 Cressier FR / Switzerland  
Phone: +41 (0)26 674 51 11  
Fax: +41 (0)26 674 54 45

Cressier, March 26<sup>th</sup>, 2018

If you are not the end user, please forward this information to the appropriate laboratory personnel

**Urgent: Field Safety Notice**  
**Ref: 002-18 / PQN 2017-010**

**Affected devices:**

Product Name	Catalog No	ID number	Lots
LISS/Coombs	004014 / 004017 / 004016 / 004015	50531	Refer to attachment 1
Coombs Anti-IgG	004023/004024/ 004025/004026/ 004027	50540	
DiaScreen	004704 / 004707 / 004706 / 004705	50571	
LISS/Coombs + Enzyme Test	004514 / 004517 / 004516 / 004515	50581	
DiaClon Type + Screen	002437 / 002431 / 002439 / 002438	50682	

Dear Customer,

This letter contains important information related to a trend review of complaints involving reagents of the ID-System.

**Description of the issue:**

We have been able to confirm an increased level of **Antibody of Undetermined Specificity (AUS)** when using some reagents of the ID-System intended for irregular antibody screening and identification: ID-Cards for Indirect Antiglobulin Testing (IAT) associated to Reagent Red Blood Cells.

AUS is defined as unexplained reactions when antibodies against red blood cell antigens have been ruled out based on non-reactivity with Reagent Red Blood Cells. For instance, the patient's sample reacts with one of the antibody screening cells (e.g ID-DiaCell or other); whereas no specificity becomes apparent when antibody identification (with ID-DiaPanel) is performed.

Our investigation indicates the main cause of this issue to be related to a raw material used in ID-Cards for IAT. We have since received and selected new lots of raw material and have initiated manufacture and supply of new batches of the above listed products. We continue to review other actions to prevent recurrence of this issue in the future.



DiaMed GmbH  
Pra Rond 23  
1785 Cressier FR / Switzerland  
Phone: +41 (0)26 674 51 11  
Fax: +41 (0)26 674 54 45

---

**Impact on the patient:**

Provided transfusion guidelines implemented within your laboratory instruct that AUS should be assigned only after all clinically significant antibodies have been ruled out during identification, (D, C, E, c, e; K, k; Fya, Fyb; Jka, Jkb; S, s, M, N; P1; Lea, Leb), the risk of reporting inaccurate final results due to this issue is remote. However, due to a potential delay in reporting final results, we have decided to communicate this information via this Field Safety Notice.

**Actions to be taken by your laboratory:**

The capacity of impacted lots of ID-Cards to detect and identify clinically significant antibodies is not affected by the issue described above. For that reason you may continue to use these products for their intended purpose. In case you would experience an AUS level impacting your ability to render final results, please consider using an alternative method such as the tube method.

We kindly ask you to complete the attached "Reply Form for End Users" (Annex I) and return it **by April 13<sup>th</sup> 2018**.

Please note that the relevant European Regulatory Agency has been advised of this FSN.

In case of questions, in the first instance, please contact our Customer Service Laboratory:

**[slabor\\_cressier@bio-rad.com](mailto:slabor_cressier@bio-rad.com)**

Our representatives are briefed to help you manage this situation.

We apologize for any inconvenience that may have been caused by this notification and we appreciate your prompt cooperation in this matter.

Yours sincerely,

Quality Assurance Representative,  
Immunohematology Division

Diane Galéa

Vice President & General Manager  
Immunohematology Division

Ann Madden



DiaMed GmbH  
 Pra Rond 23  
 1785 Cressier FR / Switzerland  
 Phone: +41 (0)26 674 51 11  
 Fax: +41 (0)26 674 54 45

**ATTACHMENT 1:** List of lot numbers where the issue of AUS might be observed

Product Name	Catalog No	IHD Lot N°	Expiry Date
LISS/Coombs	004014 / 004017 / 004016 / 004015	50531 14 xx	05.2018 - 07.2018
		50531 15 xx	06.2018 - 08.2018
		50531 16 xx	08.2018 - 10.2018
		50531 17 xx	09.2018 - 11.2018
		50531 18 xx	11.2018 - 12.2018
		50531 19 xx	11.2018 - 12.2018
		50531 20 xx	12.2018 - 02.2019
		50531 21 xx	01.2019 - 03.2019
		50531 22 xx	03.2019- 04.2019
		50531 23 xx	04.2019 - 05.2019
		50531 24 xx	04.2019 - 06.2019
		50531 25 xx	05.2019 - 06.2019
		DiaScreen	004704 / 004707 / 004706 / 004705
50571 15 xx	08.2018		
50571 16 xx	09.2018		
50571 20 xx	01.2019		
50571 21 xx	03.2019		
50571 22 xx	04.2019		
50571 23 xx	05.2019		
Coombs Anti-IgG	004023/004024/ 004025/004026/ 004027	50540 01 xx	12.2018 - 04.2019
		50540 02 xx	03.2019 - 04.2019
LISS/Coombs + Enzyme Test	004514 / 004517 / 004516 / 004515	50581 14 xx	06.2018 - 07.2018
		50581 16 xx	08.2018
		50581 19 xx	12.2018
		50581 21 xx	02.2019 - 03.2019
		50581 22 xx	03.2019
DiaClon Type + Screen	002437 / 002431 / 002439 / 002438	50682 07 xx	03.2018
		50682 09 xx	02.2018 - 03.2018
		50682 67 xx	08.2018 - 09.2018
		50682 68 xx	09.2018 - 02.2019
		50682 69 xx	11.2018 - 02.2019





DiaMed GmbH  
 Pra Rond 23  
 1785 Cressier FR / Switzerland  
 Phone: +41 (0)26 674 51 11  
 Fax: +41 (0)26 674 54 45

Annex I

Please fill out, sign and return this form to your customer Service (enter Local information) by April 13<sup>th</sup>.

**Field Safety Notice / 002-18**  
**Reply Form for End Users**

**PRODUCTS:**

Product Name	Catalog No	ID number	Lots
LISS/Coombs	004014 / 004017 / 004016 / 004015	50531	Refer to attachment 1
Coombs Anti-IgG	004023/004024/ 004025/004026/ 004027	50540	
DiaScreen	004704 / 004707 / 004706 / 004705	50571	
LISS/Coombs + Enzyme Test	004514 / 004517 / 004516 / 004515	50581	
DiaClon Type + Screen	002437 / 002431 / 002439 / 002438	50682	

**CUSTOMER INFORMATION:**

Hospital / Laboratory	
Address (Street, Postcode, Country)	
Phone Number	
Undersigning manager name	
Customer Account Number	

**STATEMENT:**

I have read and understood this Field Safety Notice, and shared the information with laboratory staff.

Date: .....

Signature: