

To all user of following systems AXIOM Sensis XP, Sensis / Sensis Lite or Sensis Vibe Combo

Product/Trade Name:	<i>see Attachment 1</i>	EU-SRN	DE-MF-000006122
Model Number:	<i>see Attachment 1</i>	E-mail	advancedtherapies-fsca.team@siemens-healthineers.com
		Date	July, 2022
		Corrective Action ID	AX033/21/S

## Customer Safety Information (CSI) for Field Safety Corrective Action

### Subject: Mounting position of the Signal Input Box (ComboBox)

Dear Customer,

We would like to inform you about a potential issue with your AXIOM Sensis XP, Sensis/Sensis Lite or Sensis Vibe Combo system and an update of your operator manual regarding chapter "Positioning the signal input box".

#### What is the issue and when does it occur?

The Sensis Signal Input Box (SIB or iSIB) is intended to be mounted underneath the patient table. This mounting position also is taken into consideration for the safety design of the Signal Input Box system and may not be appropriately covered when a different mounting position of the Signal Input Box is applied. The Sensis Signal Input Box provides hooks as a milled item on the bottom of the box. The provided hooks support the possibility of mounting the Signal Input Box on an accessory rail (e.g. at the patient table) or in some similar environment where rail mounting is possible. When the mounting position deviates from the default position there is a need to observe additional restrictions.

#### What is the impact on the operation of the system and what are the possible risks?

In case the mounting position deviates from the default position, the following scenarios are possible:

Patient touching the Signal Input Box:

This may result in touch leakage current to the patient which can exceed the allowable values of the patient touch leakage current if the Signal Input Box is mounted in a way that allows accidental contact of the patient directly to the metal housing of the Signal Input Box.

#### Siemens Healthcare GmbH

Management: Bernhard Montag, President and Chief Executive Officer;  
Darleen Caron, Jochen Schmitz

Chairman of the Supervisory Board: Ralf P. Thomas  
Registered office: Munich, Germany; Commercial Registry: Munich, HRB 213821  
WEEE-Reg.-No. DE 64872105  
SCF V12

Spillage of liquid entering the Signal Input Box:

The Signal Input Box fulfills the requirements for protection against vertically falling liquid drops. This is regardless of the mounting position. However, observe that a different mounting position, e.g. rail mounting, might allow liquids that can be spilled horizontally towards the Signal Input Box. This is because the Signal Input Box has metal grid at the bottom side, which might be exposed with a different mounting position.

**How was the issue identified and what is the root cause?**

The issue was identified internally during follow-up inspections.

No complaints or adverse events have been reported in connection with this issue.

The documentation of the expected default mounting position was inconsistent within different documents and needs to be updated in order to show more precise instructions. This affects the Data Sheet, the Operator manual, and the planning guide as well as the installation instructions might allow misinterpreting the expected mounting position and deviate from the default position in a way that might not take the additional risks into full consideration.

**Which steps have to be taken by the user to avoid the possible risks associated with this issue?**

Please follow the instructions in the attached addendum which are a supplement to the Operator Manual chapter "Positioning the signal input box".

**What actions are being taken by the manufacturer to mitigate possible risks?**

The mounting options were retested, and the Signal Input Box was released for other mountings as underneath the table mounting under the following restrictions:

- The signal input box is mounted stationary.
- The signal input box is positioned out of patient reach.
- The signal input box is protected against liquid spillage.

The affected documentation was updated, and the documentation update will be distributed to the affected customers.

**How will the corrective action be implemented?**

This letter and the attached addendum will be distributed to affected customers as update AX033/21/S and have both to be filed with the system documentation.

**What risks are there for patients who have previously been examined or treated using this system?**

There is no risk for patients who have previously been examined or treated using affected systems.

Please ensure that all users of the affected products within your organization and others who may need to be informed will receive the safety relevant information provided with this notice and will comply with the recommendations therein.


We appreciate your understanding and cooperation with this safety advisory and ask you to immediately instruct your personnel accordingly. Please ensure that this safety advisory is retained in your product related records appropriately. Please keep this information at least until the measures have been finalized.

Please forward this safety information to any other organizations that could be affected by this measure.

If the device has been sold and is therefore no longer in your possession, please forward this safety notice to the new owner. We would also request you to inform us of the identity of the device's new owner where possible.

With best regards,

Siemens Healthcare GmbH  
Business Area Advanced Therapies (AT)

 Electronically signed by: Carsten Bertram  
Reason: I am approving this document  
Date: Jul 7, 2022 15:48 GMT+2

---

Carsten Bertram  
President Advanced Therapies

 Electronically signed by: Johann Boeck  
Reason: I am approving this document  
Date: Jul 7, 2022 15:48 GMT+2

---

Johann Böck  
Person Responsible for Regulatory Compliance

Letter of July, 2022  
To all user of following systems AXIOM Sensis XP, Sensis / Sensis Lite or  
Sensis Vibe Combo



Attachment 1

<b>Product/Trade Name</b>	<b>Model No.</b>
AXIOM Sensis, Combo 64 IECG	6623974
AXIOM Sensis, Hemo LOW	6634633
AXIOM Sensis, Combo 32 IECG	6634641
AXIOM Sensis, EP 129 IECG	6634658
Sensis	10764561
Sensis Lite	10764562
Sensis Vibe Combo	11007642