

URGENT FIELD SAFETY NOTICE

RECALL AND FIELD CORRECTION TO FOCUS FORCE Intraocular Lens

Trade Mark: FOCUS FORCE

LOT number: 15071

Models: AS60125, AS60130, F260, UF60130, A100

January 12, 2017

Re: FOCUS FORCE Intraocular Lens

FSCA-identifier: 002/AS60130/17.00

Models: AS60125, AS60130, F260, UF60130, A100

Lot numbers of affected product shipped to you: 15071

Please Note: This recall is limited to LOT number 15071 only and does not affect other lots shipped to your facility.

Dear Customers,

A trade mark FOCUSFORCE intraocular lenses with a lot number of 15071, an issue has been encountered for dioptr deviation which comes from labelling process. The patient's vision can be adversely affected if the patient does not have proper lens dioptr. As a consequence of this situation, no patient has been reported to be seriously injured. However, this may require medical or surgical intervention to permanently damage the patient's vision function or structure.

The product label is shown below to assist you in identifying the product within your facility.

XXXXX

The undersign confirms that this notice has been notified the appropriate Regulatory Agency.

This action represents a product recall and we have notified the appropriate authority of this recall.

The problem which comes from labelling process; fixed. To avoid this situation, do not use the intraocular lenses with 15071 lot number in your inventory. These lenses will be replaced with new lenses. These products will be collected in the recall process.

We kindly recommend you to read this notice carefully and complete the following activities.

1. Check your stocks immediately for the 15071 LOT number.
2. Hold all products in safety position and return the existing products to PLEASE CUSTOMIZE FOR YOUR COUNTRY
3. Send this field safety notice to all interested parties within the organization.
4. If the product with this lot number is implanted in the patient; please inform Anadolu Tıp Teknolojileri A.Ş. for any adverse event.
5. Please fill out the below customer response form and send to following e mail address. pkaymak@anadolutip.com.tr as well as to: PLEASE CUSTOMIZE FOR YOUR COUNTRY


Please use the contact information below for any problems and feedback.

We kindly ask you to respond within 7 calendar days of the receipt of this notification. We thank you for your interest and we apologize for the inconvenience.

With Best Regards.

Pınar BİRTEZ

QA Manager



Anadolu Tıp Teknolojileri A.Ş.

Organize Sanayi Bölgesi No:10 Merkez,

Sivas/Türkiye

P:+90 346 218 14 18

F: +90 346 218 14 20

pkaymak@anadolutip.com.tr

