



Urgent Field Safety Notice Urgent Product Correction

Immediate Action Required

Date Issued March 18, 2025

Product

Product Description	List Number	Serial Number	UDI
CELL-DYN Ruby	08H67-01	See Attachment 1	
CELL-DYN Ruby	08H67-03		
CELL-DYN Ruby	08H67-10		
CELL-DYN Ruby	08H67-13		
CELL-DYN Ruby	04U42-84		

Explanation

Abbott would like to inform you of an issue in the CELL-DYN Ruby system where, when expired reagents are scanned or manually entered, the system will change the expiration date to current or future date.

This change occurs without notifying the user that an expired reagent is inadvertently being used. The use of expired reagents is against the guidelines of the CELL-DYN Ruby Operator’s Manual.

The CELL-DYN Ruby Operator’s Manual will be updated to provide information regarding this issue.

Impact on Patient Results

There is potential for incorrect results. To date, there has been no known impact to patient management.

Necessary Actions to be Taken by Customer

Per the CELL-DYN Ruby Operator’s Manual, do not use expired reagents as this may compromise test accuracy.

Always use reagents within their valid expiration dates for all testing procedures.

The operator must confirm the container labeling matches the data on the reagent log entry screen dialog box.

Quality Controls should be run after any reagent lot number change, maintenance, component replacement, field service action, software change, or calibration. This should be done in accordance with your laboratory’s quality control program and regulatory requirements. The quality control programs on the CELL-DYN Ruby are designed to assess precision and accuracy, identify shifts and trends, and determine the nature and cause of errors.

Complete and return the Customer Reply Form.

If you have forwarded the product listed above to other laboratories, please inform them of this Product Correction and provide to them a copy of this letter.

Please retain this letter for your laboratory records.

**Contact
Information**

If you or any of the health care providers you serve have questions regarding this information, U.S. Customers please contact Customer Service at 1-877-4ABBOTT. Customers outside the U.S., please contact your local area Customer Service.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program at <http://www.fda.gov/MedWatch/report.htm>, by phone (1-800-332-1088), or fax (1-800-FDA-0178).

If you have experienced any patient or user injury associated with this Field Action, please immediately report the event to your local area Customer Service.
