

URGENT: FIELD SAFETY NOTICE

FSCA 4593

Increased risk of false positive *Proteus* results using BioFire® FilmArray® Blood Culture Identification (BCID) Panel (Part No.: RFIT-ASY-0126 and RFIT-ASY-0127) with bioMérieux BACT/ALERT® blood culture bottles

The purpose of this letter is to inform you that BioFire Diagnostics, LLC has identified an increased risk of false positive *Proteus* results (appearing as *Enterobacteriaceae/Proteus* on the test report) when the BioFire BCID Panel is used with some BACT/ALERT® blood culture bottles (see Table 1) with expiration dates of 19SEP2020 and beyond.

Table 1. Affected media types

BACT/ALERT® Blood Culture Bottle Catalog No.	Description
410851	BACT/ALERT® FA Plus
410852	BACT/ALERT® FN Plus
410853	BACT/ALERT® PF Plus
259790	BACT/ALERT® SN

The most probable cause for this risk is the presence of an increased level of nucleic acid from non-viable *Proteus* spp. bacteria in BACT/ALERT® Blood Culture Bottles (Table 1). **The presence of non-viable organism does not compromise the intended function of the blood culture bottles (culturing viable microorganisms).** However, the BioFire BCID Panel detects nucleic acid from viable and non-viable organisms alike. Observed *Proteus* false positives are typically seen as multiple positives with the BioFire BCID Panel because a positive blood culture is a prerequisite to a BCID test.

BioFire BCID Panel product literature includes the following limitations:

- Blood culture media may contain non-viable organisms and/or nucleic acid at levels that can be detected by the BioFire BCID Panel, leading to false positive test results. Typically, these false positives will be present with one or more additional true positive results because the BioFire BCID Panel will also detect the organism that is growing in the culture bottle.
- In some cases, the Gram stain result and results of the BioFire BCID Panel may be discrepant (for example, detection of gram-positive cocci by the BioFire BCID Panel when gram-positive cocci were not observed in the Gram stain). In these cases, the BioFire BCID Panel results should be confirmed (e.g. by culture) before reporting, unless the result is concordant with other laboratory, epidemiological, or clinical findings.

Actions to be taken by customer:



- If the BioFire BCID Panel is used to test BACT/ALERT® blood culture bottles (Table 1) with an expiration date of 19SEP2020 and beyond, positive results for *Proteus* should be confirmed by another method prior to reporting the test results. This letter supersedes all previous communications regarding *Proteus* false positive results using bioMérieux blood culture bottles.
- Please return the Acknowledgment of Receipt Form accompanied with this Field Safety Notice and return it to your local bioMérieux representative.

Actions to be taken by BioFire:

- BioFire and bioMérieux teams are coordinating efforts to resolve this issue and will provide follow-up communication when this situation has been resolved.

If you have any questions or concerns, please don't hesitate to contact your local bioMérieux representative. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.

Thank you for your understanding and patience in this matter.



Wade Stevenson
Sr. Vice President of Global Marketing
BioFire Diagnostics, LLC