



Please distribute the attached customer letter.
To the Laboratory Manager
To the attention of the Healthcare center Chairman

Address
City, Date

Our reference: FSCA 5742
PJ: Annex 1

IMPORTANT:

URGENT FIELD SAFETY NOTICE

MUELLER HINTON E AGAR-Multiple references

**Antimicrobial susceptibility testing results issue/
Pseudomonas aeruginosa-Gentamicin**

Dear valued bioMérieux Customer,

This letter is intended for MUELLER HINTON E AGAR (References: 413822, 413824, 413825, 417680) users and our records indicate that your laboratory has received one of the lot numbers listed in Annex 1.

Description of the issue

Mueller Hinton E Agar is a medium for disk diffusion antimicrobial susceptibility testing and the determination of Minimal Inhibitory Concentrations (MIC) using the ETEST® method.

Following customer complaints, a performance limitation with the **combination *Pseudomonas aeruginosa* - Gentamicin on MUELLER HINTON E AGAR** has been confirmed for all the lots listed in Annex 1.

Results for ATCC 27853 *Pseudomonas aeruginosa*-Gentamicin are outside the expected quality control range (EUCAST,CLSI) using both the disc diffusion method and the ETEST® method. Diameter results are smaller than expected and MIC results are overestimated.

Impact to customer:

The potential hazards related to the product issue are a Delayed Result with an out of range quality control result (EUCAST,CLSI) for the combination ATCC 27853 *Pseudomonas aeruginosa* - Gentamicin or an Erroneous AST result, Overestimated MIC, Over-calling Resistance on a patient sample with the combination *Pseudomonas aeruginosa* - Gentamicin (CLSI).

Subsidiary name (if applicable) / Nom de la filiale (si approprié)



Required actions:

We request that you perform the following actions at this time:

- Stop using **MUELLER HINTON E AGAR - References 413822, 413824, 413825, 417680** regarding batches mentioned in Annex 1 when testing the combination ***Pseudomonas aeruginosa* – Gentamicin**.
- Complete the Acknowledgement Form in Attachment A and return it to your local bioMérieux representative to confirm receipt of this notice.

bioMérieux is committed to providing our customers with the highest quality product possible. We sincerely apologize for any inconvenience that this may have caused you. If you require additional assistance or have any questions, please contact your local bioMérieux Customer Service representative.

Thank you for your continued use of bioMérieux products,

bioMérieux, Inc.

[\[Enter Local Contact\]](#)

Subsidiary name (if applicable) / Nom de la filiale (si approprié)



Attachment A: Acknowledgement Form.

URGENT FIELD SAFETY NOTICE

FSCA#5742 - MUELLER HINTON E AGAR – Antimicrobial susceptibility testing results issue / *Pseudomonas aeruginosa*-Gentamicin

**TO BE RETURNED TO YOUR BIO-MERIEUX CUSTOMER SERVICE AT THE FOLLOWING
FAX NUMBER : XXXXXXXXX**

Name of the laboratory:

City:

Customer number:

- I acknowledge receipt of the bioMérieux letter regarding the “MUELLER HINTON E AGAR – Antimicrobial susceptibility testing results issue / *Pseudomonas aeruginosa*-Gentamicin”
- I will implement the required actions indicated in the Urgent Field Safety Notice.
- Have you encountered an impact on patients’ results, or reports of illness or injury related to the identified issue ?
 - Yes No Not applicable

DATE

SIGNATURE :

Subsidiary name (if applicable) / Nom de la filiale (si approprié)