

Customer
Hospital
City
Postal code
Country
Attn.: XXX

[ISSUE
DATE]

Field Safety Notice: AQURE System patient mix-up

Priority Level: Urgent

Dear Radiometer Customer

Background

Radiometer has become aware that there is a potential problem relating to the AQURE System that may result in patient mix-up, when connected to some specific *non-Radiometer* devices.

The problem may appear when one of the devices listed in the table below are connected to AQURE.

Devices affected:

Driver	Device	Query for patient data to/from HIS/LIS
IL	GEM 3000	Patient ID
IL	GEM Premier 3500	Patient ID
Roche	CoaguChek XS Pro/XS Plus system/PROII	Patient list By Department
Roche	Accu-Chek Inform II system	Patient list By Department
Roche	Cobas b 123 POC system	Patient ID, Sample ID
Heamonetics	TEG 6s	Patient ID
Randox	RandoxImola	Patient ID

Prerequisite:

- The device must be one of the above listed devices
- The query functionality for patient data to/from HIS/LIS (Patient ID or Patient list By Department) on the device is used via AQURE.
- The device disconnects immediately after sending the query

The patient mix-up may occur in the following scenario:

- Step 1: A user initiates a measurement on the affected device for a patient with patient id xx223 (Patient 223).
- Step 2: The user enters patient id xx223 on the affected device.
- Step 3: The affected device queries for a patient with id xx223 in AQURE.
- Step 4: The affected device disconnects from AQURE (for any reason) immediately after sending the query.
- Step 5: AQURE processes the query and places the requested patient identification data in the output buffer.
- Step 6: The affected device reconnects to AQURE.
- Step 7: A user initiates a measurement on the affected device for a different patient with patient id xx215 (Patient 215).

- Step 8: The user enters patient id xx215 on the affected device.
- Step 9: The affected device queries for a patient with id xx215 in AQUIRE.
- Step 10: AQUIRE incorrectly responds to the query for patient 215 by transmitting the patient identification data for patient 223 to the affected device.
- Step 11: If the affected device does not verify consistency between the query and the response from AQUIRE, the user will be presented to patient identification data of a wrong patient.

The patient ID numbers used above are examples only. Issue can occur with all patient IDs.

Risk for the patient

The described error may occasionally lead to serious adverse health consequences for the patient caused by patient data mix-up. The described error may, in a worst-case scenario where a critical parameter, such as K+, glucose or pO₂, is interrogated, impact upon patient management, as, based upon wrongful clinical diagnosis inferred from the erroneous parameter, a critically ill patient may be subjected to lack of vital treatment based upon the erroneous results. This may result in permanent impairment or serious injury that would require medical or surgical intervention to preclude irreversible impairment or damage.

Affected product

All AQUIRE Systems below version 2.3.5, configured with query functionality for patient data to/from HIS/LIS using Mirth or other integration engines.

Your actions

If your AQUIRE is connected to any of the above devices we request you to:

- Cease performing queries as described in the scenario above, until the AQUIRE system has been corrected by your Radiometer representative.
- Complete the Recall Response Form and return to your Radiometer representative (last page of this letter).

Solution provided by Radiometer

Your Radiometer representative will contact you to schedule a visit or a remote session. During the visit or remote session your Radiometer representative will install a patch on your AQUIRE system, which serves to correct the issue with the patient mix-up.

Your help is appreciated

If you are not the end-user of the affected product, please ensure that this letter is distributed to the final end-user.

If you have any questions, please contact your Radiometer representative.

Radiometer sincerely apologizes for the inconvenience this situation may cause you.

Best regards,
<Radiometer distributor>

Recall Response Form

Concerning:

AQURE System patient mix-up

- I have received the customer letter, and do not have any affected devices connected.
- I have received the customer letter, and informed operators of the listed devices in my facility to cease performing queries as described in the scenario with immediate effect, until the AQURE system has been corrected by my Radiometer representative.

Hospital Name:	
Your Name:	
Date:	
Signature:	
Email Address:	