



Date: December XX, 2021

Olympus reference: **QIL 154-017**

FIELD SAFETY NOTICE

RE: SUPPORTING TRAINING VIDEO ON HANDLING AND USE OF DUODENOVideoscopes TJF-Q170V/TJF-Q190V/TJF-Q290V AND SINGLE-USE DISTAL COVER MAJ-2315

Attention: Endoscopy Department, Risk Management Department

	Model Name	Serial number
OLYMPUS DUODENOVideoscope	TJF-Q170V	All
OLYMPUS EVIS EXERA III DUODENOVideoscope	TJF-Q190V	All
EVIS LUCERA ELITE DUODENOVideoscope	TJF-Q290V	All
SINGLE USE DISTAL COVER	MAJ-2315	All

Dear Healthcare Professional:

This Field Safety Notice pertains to the above-referenced Olympus endoscope models and the single-use distal cover: DUODENOVideoscope TJF-Q170V/TJF-Q190V/TJF-Q290V and DISTAL COVER MAJ-2315.

Background:

End of last year Olympus identified an increase in the number of complaints from customers reporting that mucosal tissue was found within the single-use distal cover, MAJ-2315, after withdrawing TJF-Q190V/TJF-Q290V duodenoscopes from patients.

Since January 2021 Olympus informs customers in the interest of transparency and in support of the continued safe use of the newest Olympus duodenoscope models about the above-mentioned increased number of complaints (QIL 153-016) and reminds of the need for careful handling of the referenced Duodenovideoscopes according to their IFU. Within the last months, Olympus continued to thoroughly investigate the root cause of these reported events. As a result, Olympus found the following are potential contributing factors to the observed phenomenon: distal cover with or without a crack on the tear-off line, and/or use of suction during withdrawal.

Olympus now created a supporting training video regarding the handling and use of DUODENOVideoscopes TjF-Q170V/TjF-Q190V/TjF-Q290V AND SINGLE-USE DISTAL COVER MAJ-2315. The video can be accessed by clicking on the following link:


<https://www.olympusprofed.com/gi/ercp/31651/>

As a reminder, please find hereafter the corresponding sections of the Instructions for Use and the Quick Reference Guide for the TjF-Q190V, TjF-Q290V, TjF-Q170V models and the MAJ-2315 that contain the detailed instructions regarding the correct use of the endoscope and its accessories:

- **TjF-Q170V (RC5083 03)** 3.5 Attaching accessories to the endoscope
- **TjF-Q190V (RC6454 02)** 3.5 Attaching accessories to the endoscope
- **TjF-Q290V (RC5078 04)** 3.5 Attaching accessories to the endoscope
- **MAJ-2315 (RC4852 06)** 7.3 Attaching the distal cover
- **Quick Reference Guide (RC5911)** "How to use the single use distal cover"

Action steps to be taken by the end user:

Our records indicate that your facility has purchased one or more of the affected TjF-Q170V/TjF-Q190V/TjF-Q290V duodenoscopes. Therefore, Olympus requires you to take the following actions:

1. Carefully read the content of this Field Safety Notice.
2. Ensure all personnel are completely knowledgeable and thoroughly trained on the content of this FSN, the related training video and the corresponding Instructions for Use.
Note: The most current language versions of the Instructions for Use and the Quick Reference Guide are available on the following Olympus webpage: <http://www.olympus-europa.com>
When opening the webpage select 'Medical Systems', select 'Contact & Support', click on the magnifying symbol (), select 'Instruction Manual' and search for the relevant model (e.g. 'TjF-Q190V').
3. If you have further distributed this product, identify your customers and forward them this Field Safety Notice. Please appropriately document your notification process and let us know the end-customer feedbacks accordingly.
4. Indicate on the Reply Form that you have received and understood this Field Safety Notice including the referenced training video by filling out and returning the completed enclosed Reply Form back to your local Olympus representative XXX latest by XXX.
If you need any hard copies of the relevant Instructions for Use and/or Quick Reference Guide, kindly indicate the quantity required per Model on the Reply Form.

Your National Competent Authority has been informed of this Field Safety Notice.

We fully appreciate your prompt cooperation in addressing this situation and working with us to ensure your patients are treated only with safe and effective Olympus products that you have come to rely upon. If you require additional information or on-site support, please do not hesitate to contact Olympus directly at (XXX) XXX-XXXX from Monday through Friday or by e-mail at XXX.

Sincerely,



REPLY FORM – QIL 154-017

OLYMPUS URGENT FIELD SAFETY NOTICE		
SUPPORTING TRAINING VIDEO ON HANDLING AND USE OF DUODENOVideoscopes TJF-Q170V/TJF-Q190V/TJF-Q290V AND SINGLE-USE DISTAL COVER MAJ-2315		
[Name & Address of Hospital/Medical Facility]		
[Dept/Attn]		
[Date]		
Optional: Additional Instructions for Use and/or Quick Reference Guide request Should you require hard copies of the mentioned Instructions for Use and Quick Reference Guide, please indicate in the table below behind each model the total quantity required. Your local Olympus organization will then arrange to send you the hard copies accordingly.		
Model name	Quantity of Instructions for Use required	Quantity of Quick Reference Guides required
TJF-Q170V		
TJF-Q190V		
TJF-Q290V		

I herewith acknowledge the receipt of your Field Safety Notice.

Further I confirm that I have transferred the content of the attached FSN to all affected departments on which this action has an impact. I understand the necessity of following the TJF-Q170V/TJF-Q190V/TJF-Q290V Instructions for Use carefully.

Name (Signature) _____

Name (Print) _____

Position _____

Please send your completed paper form response to XXXXX latest by XXXX.