

Date: 26:05:2023

Urgent Field Safety Notice

Intersurgical Superset Fixed Elbow Catheter Mount 22F-22M/15F ≥ 70mm-150mm

For Attention of*: All clinical staff, Managers and users of the above product.

Contact details of local representative (name, e-mail, telephone, address etc.)*

Giedrius Budrys
Customer Resolution and Relationship Manager
Intersurgical UAB
Arnioniu str 60, LT-18170 Pabrade Lithuania

Email: giedriusb@intersurgical.lt

Tel. +370 387 66611

Fax: +370 387 66622

or

This could be a distributor or local branch of the manufacturer. To be added at the appropriate stage in the different local languages

Urgent Field Safety Notice (FSN)

Intersurgical Superset Fixed Elbow Catheter Mount 22F-22M/15F ≥ 70mm-150mm

Risk addressed by FSN

| 1. Information on Affected Devices* | |
|--------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1. | 1. Device Type(s)* Catheter Mount |
| 1. | 2. Commercial name(s) Superset Fixed Elbow Catheter Mount 22F-22M/15F ≥ 70mm-150mm |
| 1. | 3. Unique Device Identifier(s) (UDI-DI) N/A |
| 1. | 4. Primary clinical purpose of device(s)* To make secure gas tight connections to other respiratory standard taper connectors and / or provide a respiratory pathway between a breathing system and a patient's airway, facemask, suction ports or monitoring ports. |
| 1. | 5. Device Model/Catalogue/part number(s)* REF: 3502000 |
| 1. | 6. Software version N/A |
| 1. | 7. Affected serial or lot number range Lot: 7220483 |
| 1. | 8. Associated devices N/A. |

| 2. Reason for Field Safety Corrective Action (FSCA)* | |
|-------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 2. | 1. Description of the product problem* We have received reports of the patient elbow disconnecting from the Superset tube in a number of products. |
| 2. | 2. Hazard giving rise to the FSCA* Whilst the security of connections should be checked during pre-use checks as per the Instructions For Use provided, if disconnection of the patient elbow was to occur in use, ventilation of the patient could be compromised. |
| 2. | 3. Probability of problem arising 1:10,000 - 1:1,000 |

| | | |
|------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------|
| 2. | 4. Predicted risk to patient/users Major risk of harm and possible occurrence. | |
| 2. | 5. Further information to help characterise the problem N/A | |
| 2. | 6. Background on Issue Intersurgical has received reports, where the patient elbow has disconnected from the Superset tube due to an insecure connection. This is a result of a process non-conformity during the assembly of the catheter mount, where the patient elbow is not fully inserted in to the Superset tube beyond the clip feature that secures it in place. | |
| 2. | 7. Other information relevant to FSCA The manufacturing process for the Catheter Mount has been investigated and the problem resolved. No other Lot numbers or products are affected. | |
| 3. Type of Action to mitigate the risk* | | |
| 3. | 1. Action To Be Taken by the User* <input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input checked="" type="checkbox"/> Destroy Device <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Follow patient management recommendations <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU) <input type="checkbox"/> Other <input type="checkbox"/> None Identify and immediately quarantine any remaining stock of the affected code and lot number listed above and do not use these devices. Please complete the Reply Form to confirm the products have been disposed of locally or to arrange collection of the devices and a credit. If you have no affected devices in stock, please confirm this using the Reply Form. Return the completed Reply Form to to giedriusb@intersurgical.it (local contact e-mail address). Please continue to report to Intersurgical any adverse events involving this product. | |
| 3. | 2. By when should the action be completed? | Immediately on receipt of this FSN and ongoing until no affected stock listed in this FSN is remaining. |
| 3. | 3. Particular considerations for: Is follow-up of patients or review of patients' previous results recommended? Not applicable. | |
| 3. | 4. Is customer Reply Required? * (If yes, form attached specifying deadline for return) | Yes |

| | | |
|-----------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| 3. | 5. Action Being Taken by the Manufacturer | |
| | <input checked="" type="checkbox"/> Product Removal <input type="checkbox"/> Software upgrade <input type="checkbox"/> Other | <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> IFU or labelling change <input type="checkbox"/> None |
| 3 | 6. By when should the action be completed? | One month of receipt of the FSN |
| 3. | 7. Is the FSN required to be communicated to the patient /lay user? | No |
| 3 | 8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet? | |
| | No | |

| | | |
|----|--------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------|
| | 4. General Information* | |
| 4. | 1. FSN Type* | New - Recall |
| 4. | 2. For updated FSN, reference number and date of previous FSN | N/A |
| 4. | 3. For Updated FSN, key new information as follows: | |
| | N/A | |
| 4. | 4. Further advice or information already expected in follow-up FSN? * | No |
| 4 | 5. If follow-up FSN expected, what is the further advice expected to relate to: | |
| | N/A | |
| 4 | 6. Anticipated timescale for follow-up FSN | N/A |
| 4. | 7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN) | |
| | a. Company Name | Intersurgical Ltd. |
| | b. Address | Crane House, Molly Millars Lane, Wokingham Berkshire, RG41 2RZ |
| | c. Website address | https://www.intersurgical.com |
| 4. | 8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. * | |
| 4. | 9. List of attachments/appendices: | Customer Reply Form |
| 4. | 10. Name/Signature | Ivan Seniut, Group Quality and Regulatory Affairs Director, Intersurgical |
| | | |

| Transmission of this Field Safety Notice | |
|-------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| | <p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback..*</p> |

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.