



VOULUNTARY FIELD SAFETY CORRECTIVE ACTION	
Description	Potential Incomplete Seal for Specific lots of Alcon Standalone Vitrectomy Consumables
Product Reference	Alcon Standalone Vitrectomy Consumables
Market Action Identifier	2025.011

August 11, 2025

Medic d.o.o.

Dear partner,

The purpose of this letter is to notify you that Alcon has initiated a Voluntary Field Safety Corrective Action for specific lots of Alcon Standalone Vitrectomy Consumables.

The following potentially affected product has been shipped to your facility:

Product #	Product Description	Affected Lot(s)
8065750957	ASSY,SHIP,CONSTELLATION VFC	179LP2
8065750957	ASSY,SHIP,CONSTELLATION VFC	17DL46

Description of the Issue

Alcon is conducting this Medical Device Field Safety Corrective Action of Alcon Standalone Vitrectomy Consumables as there is potential for some pouches within impacted lots to have an incomplete seal. Please see photos next page for examples of incomplete seal.

Due to the risk that the sterile barrier may be compromised, Alcon is recalling potentially affected lots. The use of non-sterile surgical products may increase the risk of post-operative infection, which may require additional medical and/or surgical intervention.

This event was identified internally, and to date Alcon has not received any reports of customer complaints or adverse events related to this issue.

Exhibit 1: Images of Incomplete Seal



Actions to be taken by the Customer / User

We are asking that you locate and dispose of any affected lots of Alcon Standalone Vitrectomy Consumables remaining in your inventory. To comply with this Voluntary Field Safety Corrective Action and request the replacement of any unused product, please take the following steps:

1. Review your inventory to determine if you have any unused affected product within your facility. **See table on page 1 for affected Standalone Vitrectomy Consumable lots** shipped to your location.
2. Segregate and **dispose** of any unused affected product from your inventory.
3. Call Alcon Customer Service to arrange for replacement of your affected inventory of Alcon Standalone Vitrectomy Consumables.
4. Respond to Alcon indicating your understanding of these instructions **even if you have zero (0) units remaining in inventory** by completing and returning the attached "Response Form" and returning to Alcon via email or fax.
5. Please forward this notification to all departments within your organization who may be in possession of this affected product; and any other organization to which this product may have been transferred.

Contact for Further Questions about this Medical Device Field Safety Corrective Action

In the event you have experienced adverse events or product quality issues related to this communication, please contact Alcon at qa.complaints@alcon.com.

Should you have any questions or concerns about this matter, please feel free to contact Alcon Customer service or contact your Alcon Sales Representative.

Sincerely,



Sanja Banek Babic
Regulatory Affairs Associate Director CSEEC

RESPONSE FORM

**Potential Incomplete Seal for Specific lots of
Alcon Standalone Vitrectomy Consumables
MA# 2025.011**

Medic d.o.o.

To comply with this Medical Device Field Safety Corrective Action, please take the following steps:

1. Review your inventory to determine if you have any unused affected product within your facility. See table below for affected Standalone Vitrectomy Consumable lots shipped to your location.

Product #	Product Description	Affected Lot(s)	Units Disposed

2. Segregate and **dispose** of any unused affected product from your inventory.
3. Call Alcon Customer Service to arrange for replacement of your affected inventory of Alcon Standalone Vitrectomy Consumables.
4. Respond to Alcon indicating your understanding of these instructions **even if you have zero (0) units remaining in inventory** by completing and returning this "Response Form" and returning to Alcon via email or fax.
5. Please forward this notification to all departments within your organization who may be in possession of this affected product; and any other organization to which this product may have been transferred.

Please return this Response Form via fax or email to Alcon:

Email: jelena.bjelanovic@alcon.com and sanja.banek-babic@alcon.com

Your signature below attests that you have read and understood this notification.

Signature:

Date:

Printed Name:

Title: