



FSN Ref: 2021-09 (04)
Date: 05.OCT.2021

FSCA Ref: 2021-09 (04)

Urgent Field Safety Notice
Mölnlycke® Mepilex Border Lite

For Attention of: Theatre Manager

Contact details of local representative (name, e-mail, telephone, address etc.)

Name: Local Customer Care contact will be added for each specific market

Email: XXX.XXX@molnlycke.com

Telephone: +XXXXXXXXXXXXXXXX



FSN Ref: 2021-09 (04)
Date: 05.OCT.2021

FSCA Ref: 2021-09 (04)

Urgent Field Safety Notice (FSN)
Mölnlycke® Mepilex Border Lite

1. Information on Affected Devices	
1.	1. Device Type(s) Mölnlycke® Mepilex Border Lite, 281200-01 7,5x7,5cm
1.	2. Commercial name(s) See Appendix I Product Table
1.	3. Primary clinical purpose of device(s) The clinical purpose of Mölnlycke® Mepilex Border Lite is the management of a wide range of non/low exuding wounds, such as leg and foot ulcers, pressure ulcers, surgical wounds and traumatic wounds e.g. abrasions, blisters and skin tears. Mepilex Border Lite can also be used as protection of compromised and/or fragile skin.
1.	4. Device Model/Catalogue/part number(s) See Appendix I Product Table
1.	5. Affected serial or lot number range See Appendix I Product Table

2 Reason for Field Safety Corrective Action (FSCA)	
2.	1. Description of the product problem* Mölnlycke has identified a potential safety issue. Products intended to be sterile are delivered to customers in a non-sterile state. Mölnlycke has decided to perform a Recall .
2.	2. Hazard giving rise to the FSCA* If unsterile wound products are used there is a risk of a local wound infection.

3. Type of Action to mitigate the risk	
3.	<p>1. Action To Be Taken by the User</p> <p><input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Return Device</p> <p>We need your help in ensuring that all affected products are located and that below actions are performed.</p> <p>Please follow below instructions:</p> <ol style="list-style-type: none"> 1. Identify and isolate the unused Mölnlycke® products at your facility, please see Appendix I for affected product information. 2. Fill out the Customer Reply Form or Distributor Reply Form, with quantity of identified affected products. Please sign and email/fax the Customer Reply Form or Distributor Reply Form per its instructions within 10 business days. 3. Even if you no longer have any concerned Mölnlycke® products, fill out the Customer Reply Form or Distributor Reply Form and return it back within 10 business days. Mölnlycke needs to be sure all customers are aware of the situation.



FSN Ref: 2021-09 (04)

FSCA Ref: 2021-09 (04)

Date: 05.OCT.2021

4. Mölnlycke will contact you and arrange for collection of the product(s) from your facility, as soon as you return the **Customer Reply Form or Distributor Reply Form**. Mölnlycke will issue a credit for the goods returned.
5. If you have forwarded any affected products to other healthcare institutions, please send them a copy of this **Field Safety Notice**. Make sure they act accordingly.
6. If you are a distributor, please inform your customers by sending them a copy of this **Field Safety Notice**. Make sure they act accordingly and return the **Customer Reply Form** with information collected from your end users.


We apologize for any inconvenience this will cause you, and rest assured it is our utmost intent to make this process as easy for you as possible.

In addition, Mölnlycke appreciates your help in collecting data on product complaints and/or incidents related to the concerned product. Please follow the reporting procedures established by your facility.

3.	2. Is customer Reply Required?	Yes (Within 10 business days)
----	--------------------------------	-------------------------------

FSN Ref: 2021-09 (04)
Date: 05.OCT.2021

FSCA Ref: 2021-09 (04)

4. General Information		
4.	1. FSN Type	New
4.	2. Further advice or information already expected in follow-up FSN?	No
4.	3. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Mölnlycke Health Care AB
	b. Address	Box 130 80, SE-402 52 Gothenburg, Sweden
	c. Website address	www.molnlycke.com
4.	4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.	
4.	5. List of attachments/appendices:	Appendix I Product table Appendix II Customer Reply Form
4.	6. Name/Signature	Annika Hallberg, Global Products Complaint Manager
		

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p>

FSN Ref: 2021-09 (04)

Date: 05.OCT.2021

Appendix I

FSCA Ref: 2021-09 (04)

Product table

MATERIAL	MATERIAL DESCRIPTION	BATCH
281200-01	Mepilex Border Lite 7,5x7,5cm	21296906