



AMANNGIRRBACH

Amann Girschbach AG | Herrschaftswiesen 1 | 6842 Koblach | Austria

Customer name & address

Koblach, 26.04.2021

Fon +43 5523 62333-260
Fax +43 5523 62333-249

Field Safety Notice for Zolid FX Multilayer and Zolid gen-x

Dear Sir or Madam,

Today, we want to inform you about the possibility of a contamination of our blanks in the following production batches of Zolid FX Multilayer and Zolid gen-x.

In these batches, there is a potential risk of foreign particles introduced during the production process that could affect the aesthetics as well as mechanical properties of the final product.

To date, no complaints related to the above have been reported to us. This safety information is a voluntary, precautionary measure to ensure the high quality and safety standards of Amann Girschbach products.

We therefore ask you to return the products you have in stock with the batch numbers mentioned below to Amann Girschbach AG. Please enter the number of blanks you will be returning in the table below.

Article number	Article	Batch/LOT	Number returned
761778	Zolid FX ML 0/A1 98X20N	2103000	
761810	Zolid FX ML D2/D3 98X20N	2103001	
766945	Zolid gen-x D3 98X16	2103000	

We have already corrected this defect internally and assure you a perfect replacement.

Thank you in advance for your support.

Wolfgang Reim, CEO

Eelco De Groot, COO

This letter has been typed and is therefore valid without signature.