

VOLUNTARY MEDICAL DEVICE FIELD CORRECTION	
Description	Increased Complaints for Knife Sharpness
Relevant Products	ClearCut® Dual Bevel 1.2mm Sideport, ClearCut® Safety Dual Bevel 1.2mm Sideport, ClearCut® Safety Dual Bevel 1.0mm Sideport and 20G MVR+ V-Lance® Ophthalmic knives
Action Identifier	2024.007

May 07, 2024

Medic d.o.o.

Dear Healthcare Professional,

Alcon has detected an increase in complaint reports related to sharpness for certain ClearCut® Dual Bevel Sideport and A-OK® Corneal/Scleral V-Lance® Ophthalmic knives. Our records indicate that potentially affected product has been shipped to your site, either as sterile standalone knives, or within an Alcon Custom Pak®.

This event was identified through our customer complaints process. Alcon has received reports of adverse events related to this event.

Products in scope:

Material #	Product Name
8065921541	CLEARCUT Dual Bevel 1.2mm Angled Sideport Knife
8065771541	CLEARCUT Safety Dual Bevel 1.2mm Angled Sideport
8065771540	CLEARCUT Safety Dual Bevel 1.0mm Angled Sideport
8065912001	20 Gauge MVR+ Blade V-Lance

Potential patient impact:

If a knife does not meet expectations for sharpness, the need for increased penetration force may be experienced during the initial stab wound. If you perceive that your knife does not meet your sharpness expectation, please cease use of the specific knife, and replace the knife with a new one.

Using extra penetration force during an anterior segment ophthalmic surgical procedure has the potential to cause intraoperative complications such as corneal abrasion, Descemet's detachment, or wound leakage. Extra penetration force during vitreo-retinal surgery could lead to ciliary body detachment, vitreous base traction, or retinal tear in the worst-case scenario.

Actions to be taken by the Customer / User:

If a knife does not perform as expected, or if you feel it does not meet your expectation for sharpness, discontinue use of that specific knife and replace the knife with a new one to avoid potential complications.

To acknowledge your receipt of this Voluntary Medical Device Correction notification, please take the following steps:

1. Forward this notification to all departments or organizations using Alcon Ophthalmic Knives.
2. Follow the risk mitigation precautions provided in this notice when using identified catalogue numbers of ophthalmic knives.
3. Please complete the attached "*Response Form*" indicating your understanding of the included instructions and **return the attached "*Response Form*" via email or fax to Alcon.**

Alcon has reported this issue to Health Authorities in accordance with applicable regulations.

In the event you have experienced adverse events or product quality issues related to this communication, please contact Alcon via web (<https://notifeye.alcon.com>), by email (msus.safety@alcon.com) or by phone (+42 0775884204).

Adverse events or quality problems experienced with the use of this product may also be reported to the local Health Authority.

Should you have any questions or concerns about this matter or need help finding a replacement or substitute sterile standalone ophthalmic surgical knife, please call Alcon Customer Service or contact your Alcon Sales Representative.

Sincerely,



Sanja Banek Babic
Alcon Farmaceutika d.o.o.